

राष्ट्रीय केला अनुसंधान केन्द्र (भारतीय कृषि अनुसंधान परिषद) तोगामलै रोड, तायनुर पोस्ट, तिरुचिरापिछ्ठ. - ६२० १०२. तमिलनाडु

National Research Centre for Banana (Indian Council of Agricultural Research) Thogamalai Road, Thayanur P.O. Tiruchirapalli - 620 102, Tamil Nadu, India

F.No.25(1)/2016-SP/ 3541 - 3549

Dated: 02.05.2016

To

Sub

Supply of Lysimeters (2 Nos.) – reg.

Ref

This office quotation invitation of even no. dated 11.04.2016.

Sir/s,

With reference to the subject cited, I am directed to inform you that the last date for submission of quotation for supply and installation of "LYSIMETERS (2 NOS.)" as detailed below is extended up to 23.05.2016 (Monday). Hence, you are once again requested to submit your competitive quotation for supplying these items immediately to the Director, ICAR-NRC for Banana, Trichy.

S.No.	Name of the article	Qty. Reqd.
1.	LYSIMETER (Digital Gravimetric Weighing Lysimeter)	.2 Nos.
•	Size: 2 m Diameter x 1 m Depth – Cylindrical – Two numbers to be	
	installed in the field with following specifications:-	
	Building/Tank Material: Stainless Steel/Mild Steel Minimum 6 mm	
	Thickness (Welded properly without any leakage) with tapper at Bottom	
	for Leached Collection.	
	Weighing Equipment: Load Cells (weighing sensors)—Precision — Output	1 1 1
	to be connected to a data logger or a Digital Panel meter	
	Soil Capacity: Up to 4000 kg.	
	Seepage water Determination: Suction Cups evacuated through motorized	
	pump.	
	Probe to be installed: Electronic Tensiometer, FDR Soil Moisture Sensor,	
	Leaf Wetness Sensor, Soil Temp Probe Two Depths Each and one	
	reference each.	
	Data recording: An hourly interval	*
	Battery recharge through solar panel regularly / weekly	1.
12	Data retrieval port from data logger through / pen drive to analyse in	2 0
	computer.	
	Optional output through GPRS modem should be quoted.	
	Outer Tank: Fabricated from this gauge Stainless Steel/Mild steel to be	
	painted	
	Complete installation and training should be provided to the user free of	
	cost.	
	Other optional / essentials accessories to be quoted	
	Lysimeter's Data logger specification: Enclosed as Annexure I	
\	TWO YEARS FREE WARRANTY.	

Nes (5)

Note:

- 1. The quotation should be sent ONLY BY SPEED POST.
- 2. Necessary leaflet / broucher of the model quoted should be enclosed, failing which, the quotation/offer will be rejected.
- 3. If you have already submitted your offer, this reminder may be ignored.

The right to accept or to reject all or any of the quotation in part or full is reserved.

ASSISTANT ADMINISTRATIVE OFFICER FOR DIRECTOR

Copy to:

1. Chairman (SPAC)

2. Member Secretary (SPAC)

3. The Chairman (PMEC) – for uploading in NRCB Website.

4. The Nodal Officer (CPPP) - for uploading in CPP Portal.

5. Pvt. Secry. to Director.