



भाकृअनुप-राष्ट्रीय केला अनुसंधान केंद्र
(भारतीय कृषि अनुसंधान परिषद)
तोगमलै रोड, थायनूर पोस्ट,
तिरुच्चिरापल्लि, -६२० १०२, तमिलनाडु

ICAR - National Research Centre for Banana
(Indian Council of Agricultural Research)
Thogamalai Road, Thayanur P.O.
Tiruchirapalli - 620 102, Tamil Nadu, India



F.No.28(1)/2016-SP/ 1507 - 1519
To

Dated : 24.10.2016

Sir/s,

Quotations are invited by the Director, ICAR - NRC for Banana, Trichy, for the supply of the following items as per specifications given below.

Name of the item	Qty.
Laminar Flow chamber (as per specification enclosed)	1 No.

1. Quotations should be inclusive of all charges towards packing, forwarding, Insurance, freight etc. Quotations exclusive of packing and freight charges etc. should indicate the amount that shall be charged on this account.
2. Complete descriptions, specifications and make of the goods should be given in the quotations. Necessary literature and pamphlets if any also be sent along with the quotation. Rate per unit should also be given precisely and delivery period clearly mentioned.
3. **THE FIRM SHOULD INVARIABLY QUOTE THE TIN, CST NO. ETC. ON THE BODY OF THE LETTERHEAD IN WHICH THE QUOTATION IS SENT, IF NOT QUOTATION WILL BE REJECTED.**
4. The quotation should be F.O.R. destination and should be kept open for 180 days from the date of quotations. Delivery at the Institute premises at NRCB Office Cum Laboratory Building, Near Thayanur Santhai, Thogamalai Road, Tiruchirappalli-620102/NRCB Research Farm Podavur Village, Inampuliyur Post, Via Kuzhumani, Tiruchirapalli - 639-103 (Tamil Nadu) is preferable.
5. The quotations should be addressed in the name of **"DIRECTOR, NATIONAL RESEARCH CENTRE FOR BANANA, Near Thayanur Santhai, Thayanur Post, Thogamalai Road, Tiruchirapalli - 620 102 (T.N.)"**, and should reach **LATEST by 08.11.2016 (Tuesday) by 3.00PM**, in a sealed cover superscribing **"Quotations for the supply of LAMINAR FLOW CHAMBER"**

The right to accept or to reject all or any of the quotation in part or full is reserved.

NOTE : You are requested to send your quotations through "SPEED POST" only and not through "Courier service".

Yours faithfully,

Senior Administrative Officer

Copy to:-

1. Chairman (SPAC)
2. The Chairman (PMEC) – for uploading in NRCB Website.
3. The Nodal Officer (CPPP) - for uploading in CPP Portal.
4. PS to Director

Phone : 0431-2618125 | Fax : 0431-2618125

Email : director.nrcb@icar.gov.in | directornrcb@gmail.com

Skype : nrcb.director | website : www.nrcb.res.in



(2)

Laminar Flow Specifications:

(Double deater)

Blow type horizontal laminar air flow containing HEPA / ULPA filter.

Nylon pre filter

Filter pore size should be less than 0.2 μ M

UV facilitated and cool white light required is 2500 lux intensity

All side of the chamber is UV block materials

Work bench made of SS304 or higher

Door of the chamber is manual opening and folded UV resistant transparent walls and they should be guaranteed for crack proof upon UV exposure.

No noise during operation

Control of air flow by two / three speed without noise

Two / Three 5Amp plug points in the need to be provided

Equipped with glass bead sterilizer up to 200°C temperature

Servo controlled automatic voltage stabilizer (5 KVA) to support this machine.

Cushion chairs

Size of the Cabinet:

Horizontal - Required size of the working chamber 6 (l) X 3(h) X 3(w) ft.